



ELECTRONIC BALANCES AND SCALES

SUBSTANCE – PAPE CLEANING GEL

Below information on determination of humidity content for above specified substance is a supplement of traditional drying method. RADWAG technical support department hopes, that this datasheet will be helpful for the tests performed with application of RADWAG moisture analyzers.

DRYING PARAMETERS – TRADITIONAL METHOD

Substance description:	Pipe cleaning gel
Reference /traditional/ method parameters:	105 °C / 4 hours.
Dry mass determined with traditional method:	12,58 %
Standard deviation	0,057 %



DRYING PARAMETERS - MOISTURE ANALYZER SERIES MAC 50

Equipment necessary for drying process	Disposable aluminum pans
Sample preparation	Pour adequate amount of gel onto a glass fibre filter
Sample size	~ 1 g
Drying profile	Standard
Drying temperature	125 °C
Auto switch-off criterion	Automatic 1 [Auto 1 - 1mg/10sec; Auto 2 - 1mg/25sec; Auto 3 - 1mg/60sec; Auto 4 - 1mg/90sec; Auto 5 - 1mg/120sec]
Average measurement time	~ 9 minutes
Determined humidity content /average value/	12,47 %
Standard deviation	0,040 %

© 2008 RADWAG. All Rights Reserved.

All copyrights to the materials contained in the publication belong to RADWAG and may not be copied, processed or used for commercial purposes without their knowledge or consent RADWAG.